



Norsonic AS  
Gunnarsbråtan 2  
N-3409 Tranby  
Norway  
Tel: +47 3285 8900  
info@norsonic.com  
www.norsonic.com

## **Declaration of Conformity**

We, Norsonic AS, Gunnarsbråtan 2, Tranby, Norway, declare under our sole responsibility that the product:

### **Airflow Resistance Measurement System Nor1517A**

to which this declaration relates, is in conformity with the following standards or other normative documents:

Method	:	ISO 9053-2: 2020
Safety	:	EN 61010-1:2010/A1:2019 Portable equipment and pollution category 2.
EMC	:	EN 61000-6-3:2007/A1:2011, EN 61000-6-2:2019
RoHS	:	Directive 2011/65/EU, 2015/863/EU
WEEE	:	Directive 2012/19/EU

This product has been manufactured in compliance with the provisions of the relevant internal Norsonic production standards. All our products are tested individually before they leave the factory.

Nor1517A measures the specific airflow resistance – and related quantities – by the applying the alternating airflow method (Method B) according to ISO 9053: Acoustics – Materials for acoustical applications – Determinations of airflow resistance.

This Declaration of Conformity does not affect our warranty obligations.

Tranby, September 2021

  
Jens Petter Ringvold  
Quality Manager