



Norsonic AS
Gunnarsbråtan 2
N-3409 Tranby
Norway
Tel: +4732858900
info@norsonic.com
www.norsonic.com

Declaration of Conformity

We, Norsonic AS, Gunnarsbråtan 2, N-3409 Tranby, Norway, declare under our sole responsibility that the product:

Sound Level Meter / Real Time Analyser Nor140 Nor1209 / Nor1209A preamplifier


to which this declaration relates, is in conformity with the following standards or other normative documents

| | | |
|-----------|---|--|
| Standards | : | EN/IEC 61672-1:2013 Class 1, EN/IEC 61260 class 1:2014 , ANSI S1.4 2014 Class 1, ANSI S1.43 1997 class 1, ANSI S1.11 2004 class 1 |
| Safety | : | EN 61010-1:2010/A1:2019 Portable equipment and pollution category 2 |
| EMC | : | EN/IEC 61672-1:2013 Class 1 |
| RoHS | : | Directive 2011/65/EU, 2015/863/EU |
| WEEE | : | Directive 2012/19/EU |

This product has been manufactured in compliance with the provisions of the relevant internal Norsonic production standards. All our products are tested individually before they leave the factory. Calibrated equipment – traceable to national and international standards – has been used to carry out these tests.

This Declaration of Conformity does not affect our warranty obligations.

Tranby, September 2021


Jens Petter Ringvold
Chief Engineer